anresco LABORATORIES since 1943

Certificate of Analysis

November 18, 2025

KATS BOTANICALS
750 Hickory Industrial Dr.
Old Hickory, TN 37138

Order No. 581643
Sample No. 1360835

SAMPLE INFORMATION

Description	M300
Lot Number	5460
Received	November 12, 2025
Serving Size (g)	1

ANALYTICAL RESULTS

Analysis Residual Solvent Screen ✔ Pass
Method USP <467>
Analysis Date November 12, 2025 to November 18, 2025

Analyte	LOD/LOQ (ppm)	Findings (ppm)	Limit (ppm)	Status
1,2-Dichloroethane	0.2/0.5	ND	5	Pass
Acetone	67/200	ND	5000	Pass
Acetonitrile	67/200	ND	410	Pass
Benzene	0.2/0.5	ND	2	Pass
n-Butane	67/200	ND	-	-
Chloroform	0.2/0.5	ND	60	Pass
Ethanol	67/200	ND	5000	Pass
Ethyl acetate	67/200	ND	5000	Pass
Ethyl ether	67/200	ND	5000	Pass
Ethylene oxide	0.2/0.5	ND	10	Pass
n-Heptane	67/200	ND	5000	Pass
n-Hexane	67/200	ND	290	Pass
Isopropyl alcohol	67/200	ND	5000	Pass
Methanol	67/200	ND	3000	Pass
Methylene chloride	0.2/0.5	ND	600	Pass
n-Pentane	67/200	ND	5000	Pass
Propane	67/200	ND	-	-
Toluene	67/200	ND	890	Pass
Total xylenes (ortho-, meta-, para-)	67/200	ND	2170	Pass
Trichloroethylene	0.2/0.5	ND	80	Pass

NT = Not Tested
ND = None Detected
LOD = Limit of Detection
LOQ = Limit of Quantitation

Reported by
Anresco, Inc.

ANRESCO, INCORPORATED OCTOBER 1 1978 CALIFORNIA

Alma Dorantes
Analyst I

November 18, 2025

If there are any questions with this report, please contact "compliance@anresco.com".

This document is intended only for the use of the party to whom it is addressed and may contain information that is privileged, confidential or protected from disclosure under applicable law. Sampling was not performed by the laboratory. Results apply only to the sample(s) as received. Anresco Laboratories is ISO/IEC 17025:2017 Accredited by ANAB - Certificate Number AT-1551.